

Underwriters Laboratories (UL LLC) Safety Certification Report



Model: TR36Mxxx (where xxx = 050, 090, 120, 135, 150, 180, 240, 360 and 480 for denoting output rating)

Device Description: Switching Power Adapter

Applicant: CINCON ELECTRONICS CO LTD
8-1 FU KUNG RD, FU HSING PARK, FU HSING HSIANG,
CHANGHUA HSIEN, 506 TAIWAN

Manufacturer: Same as Applicant

Manufacturing Facility(ies): DONGGUAN CINCON ELECTRONICS LTD
66 YUANYING RD, LIAOBU TOWN,
DONGGUAN
GUANGDONG 523412 CHINA

CINCON ELECTRONICS CO LTD
8-1 FU KUNG RD, FU HSING PARK, FU HSING HSIANG,
CHANGHUA HSIEN, 506, TAIWAN

Report No.: E252331-D1024-1/A0/C0-UL

Report (Re)Issue Date: 2023-12-25

Base Standard(s): ANSI/AAMI ES60601-1:2005 & A1:2012 & A2:2021 (consolidated version) –
revision date: 2022/02/01
CAN/CSA-C22.2 No. 60601-1:14 (IEC 60601-1: 2005+A1:2012+A2:2020,
MOD) (consolidated version) – revision date: 2022/July

Additional Standards: IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020; ANSI/AAMI
HA60601-1-11:2015/A1:2021–Revision date: 2022/02/18; CSA C22.2 NO.
60601-1-11:15 (R2020) + A1:21–Revision date: 2022/06

Report Types: This report consists of the following report types:

- [Yes] US Certification (UL Recognition)
- [Yes] CAN Certification (cUL Recognition)

This report covers the Safety evaluation of the referenced model(s) according to the standard(s) specified above.